

Republic of the Philippines
Department of the Interior and Local Government
NATIONAL POLICE COMMISSION
Quezon City

RESOLUTION NO. 2017-166

**APPROVING THE PNP-UESB RESOLUTION NO. 2016-101 DATED AUGUST 4, 2016
THAT PRESCRIBES THE MINIMUM STANDARD SPECIFICATIONS FOR
INTENSIVE CARE UNIT (ICU) BED**

WHEREAS, Section 14, paragraphs (i) and (m) of Republic Act No. 6975, as amended by Section 5, paragraph (a), sub-paragraphs (5) and (8) of the Republic Act No. 8551, empowers the Commission to “approve or modify plans and programs on education and training, logistical requirements, communications, records, information system, crime laboratory, crime prevention and crime reporting as well as prescribe minimum standards for arms, equipment and uniforms and, after consultation with the Philippine Heraldry Commission, for insignia of ranks, awards and medals of honor”;

WHEREAS, pursuant to the aforecited legal prescription, the PNP through its Uniform and Equipment Standardization Board (UESB) Resolution No. 2016-101 dated August 4, 2016, recommended to the Commission for approval the proposed minimum standard specifications for Intensive Care Unit (ICU) Bed;

WHEREAS, the UESB proposed specification refers to description and parameter of an Intensive Care Unit (ICU) Bed purposely to enhance the medical capability of PNP General Hospital in providing sophisticated medical equipment to confined patients suffering from severe illness;

WHEREAS, the PNP UESB certified the proposed minimum standard specifications for Intensive Care Unit (ICU) Bed is not a tailored-fit specification to any particular brand/model of product in the market;

WHEREAS, after thorough review and evaluation of the UESB resolution on the subject proposed minimum standard specifications, the Commission finds the same in order and conforms with the requirements of Republic Act No. 6975, as amended by Republic Act No. 8551 and its Implementing Rules and Regulations (IRR), all existing policies and issuances of the Commission to meet the requirements of the PNP;

NOW THEREFORE, the Commission, in the exercise of its statutory power to administer and control the PNP, HAS RESOLVED TO APPROVE, AS IT HEREBY APPROVES, the PNP-UESB Resolution No. 2016-101 dated August 4, 2016, to prescribe the minimum standard specifications for Intensive Care Unit (ICU) Bed, with slight modifications, to read as follows:

I. TECHNICAL SPECIFICATIONS


- | | |
|-----------------------|--|
| 1. Overall Bed Length | : a) 93” minimum for extended
b) 81” minimum for retracted |
| 2. Overall Bed Width | : a) 40” minimum for siderails up
b) 36” minimum for siderails down |
| 3. Overall Bed Height | : a) Low: Down to 18”
b) High: Not more than 43” |
| 4. Caster Diameter: | : 5” to 6” |


- 5. Trendelenburg/ : 12°/12°
Reversed Trendelenburg
- 6. Weight Capacity : 227kg
- 7. Power : 220VAC

II. MUST PASS THE PNP TEST AND EVALUATION


APPROVED this 8th day of MARCH 2017 in Quezon City, Metro Manila, Philippines.



ISMAEL D. SUENO
Chairman, NAPOLCOM




ATTY. ROGELIO T. CASURAO
Commissioner
Vice Chairman and Executive Officer



ZENONIDA F. BROSAS
Commissioner


~~JOB M. MANGENTE~~
Commissioner


FELIZARDO M. SERAPIO, JR.
Commissioner


RONALD M. DELA ROSA
Ex-Officio Commissioner

Attested by:


EVVIRA M. BAUTISTA
Acting Chief Secretariat