

Republic of the Philippines
Department of the Interior & Local Government
NATIONAL POLICE COMMISSION
DILG-NAPOLCOM Center, NAPOLCOM Building
EDSA Corner Quezon Avenue, West Triangle,
Quezon City 1104

RESOLUTION NO. 2017-346

APPROVING THE PNP-UESB RESOLUTION NO. 2017-007 DATED FEBRUARY 28,
2017 THAT PRESCRIBES THE MINIMUM STANDARD SPECIFICATIONS FOR
SURGICAL SUCTION MACHINE

WHEREAS, Section 14, paragraphs (i) and (m) of Republic Act No. 6975, as amended by Section 5, paragraph (a), sub-paragraphs (5) and (8) of the Republic Act No. 8551, empowers the Commission to “approve or modify plans and programs on education and training, logistical requirements, communications, records, information system, crime laboratory, crime prevention and crime reporting as well as prescribe minimum standards for arms, equipment and uniforms and, after consultation with the Philippine Heraldry Commission, for insignia of ranks, awards and medals of honor”;

WHEREAS, pursuant to the aforecited legal prescription, the PNP Uniform and Equipment Standardization Board (UESB) through Resolution No. 2017-007, dated February 28, 2017, had proposed the minimum standard specifications for Surgical Suction Machine, for consideration and approval by the Commission;

WHEREAS, the minimum standard specifications for Surgical Suction Machine refers to the description of equipment that removes excess secretions from the mouth and throat (oropharynx), from the nose and throat (nasopharynx), and from the windpipe (trachea);

WHEREAS, the PNP issued a certification that the minimum standard specifications for Surgical Suction Machine is not tailored-fit to any specific brand of product in the market;

WHEREAS, after thorough review and evaluation, the Commission finds the proposed minimum standard specifications for Surgical Suction Machine in conformity with the requirement of Republic Act No. 6975, as amended by Republic Act No. 8551, its Implementing Rules and Regulations (IRR), and all existing rules and regulations to meet the requirements of the PNP;

NOW THEREFORE, the Commission, in the exercise of its constitutional power to administer and control the PNP, HAS RESOLVED TO PRESCRIBE, AS IT HEREBY PRESCRIBES, the minimum standard specifications for Surgical Suction Machine, to read as follows:


- I. Technical Specification
 - a) Maximum Vacuum : 720mmHg
 - b) Maximum Flow Rate : 60L/min
 - c) Voltage : 220V (50-60Hz)

- d) Number of Canister : Two (2) suction bottles
- e) Canister's Capacity : Four (4) liters
- f) Mobility/Portability : Four (4) casters with brake

II. Must pass the PNP Test and Evaluation.


APPROVED this 23rd day of JUNE, 2017, in Quezon City, Metro Manila, Philippines.


CATALINO S. CUY
Officer-in-Charge, DILG


ATTY. ROGELIO T. CASURA
Commissioner
Vice Chairman & Executive Officer



FELIZARDO M. SERAPIO, JR.
Commissioner


JOB M. MANGENTE
Commissioner


ZENONIDA F. BROSAS
Commissioner


RONALD M. DELA ROSA
Ex-Officio Commissioner

Attested by:


ELVIRA M. BAUTISTA
Acting Chief, Secretariat